opening increased by 28 mm with satisfactory lateral excursions of the mandible. The patient was pleased with her progress and no operative treatment was thought to be necessary. She had been advised to continue with the same self physiotherapeutic measures and will remain under review.

Coronoidectomy in order to release the temporalis muscle attachment from the mandible has been the preferred method of treatment for this condition, as temporalis myotomy is likely to produce only a limited period of improvement before fibrosis recommences at the site of myotomy. Coronoidectomy has been carried out in the cases previously reported.¹⁻⁴ In the case of Sanders *et al*² a bilateral procedure was carried out in order to release fibrosis which occurred following a transcoronal incision and bifrontal craniotomy for excision of an olfactory meningioma.

In the case reported here, the patient was seen in the Maxillofacial Surgery Department about a year after the neurosurgical procedure with significant limitation to mandibular opening. However, institution of active jaw exercise utilising a trismus screw was seen to have produced significant improvement in condition. It is suggested that this condition be initially treated by active jaw opening exercises, if possible in the early neurosurgical postoperative period. By doing so, surgical correction of the disability may be avoided.

MYM ZAFARULLA

Dept. of Oral and Maxillofacial Surgery,
Bolton District General Hospital,
Bolton,
Lancashire, BL4 0JR,
UK

References

- ¹ Khosla VM. Pseudo ankylosis of mandible produced by temporal muscle fibrosis: report of a case. *J Oral Surg* 1970;28:521-2.
- ² Sanders B, Thorpe W, Kallal R. Pseudo-ankylosis of the mandible secondarily to transcoronal neurosurgical procedure. *J Oral Surg* 1974;32:909-11.
- ³ Kwapis BW, Dyer MH, Edward Sville I. Pseudo ankylosis following a temporomandibular joint following craniotomy. *J Oral Surg* 1974;32:912-4.
- ⁴ Summers L. False ankylosis of the temporomandibular joint following craniotomy. *Brit J Oral Surg* 1980;18:138-140.

Accepted 16 February 1985

Matters arising

Measurement of thermal thresholds

Sir: I read the papers 'An improved automated method for the measurement of thermal thresholds. 1. normal subjects. ... 2. patients with peripheral neuropathy',^{1,2} with great interest, since quite independently, we also have been developing a microcomputer-based thermal testing system.^{3,4} By placing the power supply, signal convertors, thermometer and other meters in a single peripheral device and using a portable computer, the system we have developed is transportable, but operates on very similar principles to that described by Jamal *et al*.^{1,2}

The results obtained by these authors are impressive. There are, however two points arising from the discussion, which should be considered further.

Spatial summation is an important determinant of thermal sensitivity, particularly for warming, and values obtained for threshold are therefore very dependent on contact area.^{5,6} Our thermode has a contact area of 7.5 cm² and is sufficiently small to be used on the face and the dorsum of the hand. With this thermode we are obtaining higher values for threshold measurements³ than those given by Jamal *et al*, who used a 12.5 cm² thermode.¹ Dyck *et al* use a 3.5 cm² thermode and give results from the dorsum of the foot for 303 healthy subjects.⁴ The details of thermode size and the number of patients studied are relevant to the comparison made by Jamal *et al* between their results and those of Dyck *et al* for inter-subject variation. One would expect lower values from a 12.5 cm² contact area thermode and the absolute range of thresholds cannot be entirely unrelated to the number of subjects in the selected age bands, studied by each group. By using a relatively large thermode, Jamal *et al* may have lost the discrimination necessary to demonstrate the variations in regional sensitivity, which other workers have found.⁵⁻⁸

To present their data on intra-individual variation, Jamal *et al* take the mean value of thresholds from all subjects together, measured on one or more occasion and give the maximum differences between these means, with the difference expressed as a percentage of the first observation.¹ Fagius and Wahren⁹ presented their data for intra-individual variability as the range of differences found with paired observations in individuals, expressing the differ-

ence as a percentage of the first measurement. It was therefore incorrect of Jamal *et al* to use the figure of 150% for intra-individual variation from the work of Fagius and Wahren⁹ in direct comparison with theirs of 5%,¹ since the two studies calculated variability in a different way.

CLARE J FOWLER

The Rita Lila Weston Institute of
Neurological Studies
The Middlesex Hospital Medical School
London W1N 8AA, UK

References

- ¹ Jamal GA, Hansen S, Weir AI, Ballantyne JP. An improved automated method for the measurement of thermal thresholds. 1. Normal subjects. *J Neurol Neurosurg Psychiatry* 1985;48:354-60.
- ² Jamal GA, Weir AI, Hansen S, Ballantyne JP. An improved automated method for the measurement of thermal thresholds. 2. Patients with peripheral neuropathy. *J Neurol Neurosurg Psychiatry* 1985;48:361-6.
- ³ Fowler CJ, Burns D, Howe N. A system for measuring thresholds for hot and cold sensation. *Environmental Health* 1985;3.
- ⁴ Fowler CJ, Burns D, Howe N, Harrison MG, Le Quesne PM. A system for measuring thresholds for hot and cold sensation. Proceedings of ICECN 1985, *Electroencephalogr Clin Neurophysiol*, in press, 1986.
- ⁵ Kenshalo DR. Correlations of temperature sensitivity in man and monkey, a first approximation. In: *Sensory functions of the skin of primates*. Zotterman Y. ed. Oxford: Pergamon Press. 1976:305-29.
- ⁶ Stevens JC, Marks LE, Simonson DC. Regional sensitivity and spatial summation in the warmth sense. *Physiol Behaviour* 1974;13:825-36.
- ⁷ Dyck PJ, Karnes J, O'Brian PC, Zimmerman IR. Detection thresholds of cutaneous sensation in humans. In: *Peripheral Neuropathy*. Dyck, Thomas, Lambert and Bunge, eds. Philadelphia: Saunders. 1984:1103-38.
- ⁸ Fruhstorfer H, Linblom U, Schmidt WG. Method for quantitative estimation of thermal thresholds in patients. *J Neuro Neurosurg Psychiatry* 1976;39:1071-5.
- ⁹ Fagius J, Wahren LK. Variability of sensory threshold determination in clinical use. *J Neurol Sci* 1981;51:11-27.

Jamal *et al* reply

We have read Dr Fowler's comments with interest. Dr Fowler suggests that the differences in results obtained by ourselves compared with Dyck *et al*¹ are a consequence of thermode size. We agree that spatial summation determined in part by thermode